

**Grievance Meeting ADP Scheme "Program for Revamping of Emergencies at Teaching Hospitals in Punjab (Sheikh Zayed Hospital, R.Y. Khan)
Financial Year 2022-23 Held On 06-10-2022**

Grievance submitted by:

M/s Eastern Medical Technology Services, Lahore

Sr. No	Item Name	Model	Make	Status of Technical Committee	Observations raised by Technical Committee	Reply by the firm	Review by the Technical Member	Decision of the Grievance Committee	Current Status
1	Electrohydraulic OT Table	504325	Medifa	Non Responsive	1. Electric Height adjustment is 780 to 1000 mm instead of 730 to 1000 mm or more. 2. Electric lateral tilt 20° and -20° instead of 30° degree and -30° degree or better.	Reply submitted by firm	Objection Removed	Grievance Accepted, Objection removed	Responsive
2	Pendant (Anesthesia)	BMGPMMP	Brandon Medical	Non Responsive	2 Years Performance/Experience certificate was not attached.	Reply submitted by firm	Objection Stand	Objection Stand	Non-Responsive
3	Pendant (Surgical)	BMGMBP	Brandon Medical	Non Responsive	2 Years Performance/Experience certificate was not attached.	Reply submitted by firm	Objection Stand	Objection Stand	Non-Responsive

Grievance submitted by:


M/s G.MED Pvt.Ltd, Lahore


Sr. No	Item Name	Model	Make	Status	Observations raised by Technical Committee	Reply by the firm	Review by the Technical Member	Decision of the Committee	Current Status
1	Pendant (Anesthesia)	PGD-QJ-1	Pegisdan	Non Responsive	Performance certificate is issued by the Engineer instead Head of the Institute.	Reply submitted by firm	Objection Stand	Objection Stand	Non-Responsive
2	Pendant (Surgical)	PGD-QJ-1	Pegisdan	Non Responsive	Performance certificate is issued by the Engineer instead Head of the Institute	Reply submitted by firm	Objection Stand	Objection Stand	Non-Responsive


Grievance submitted by:


M/s Radiant Medical, Lahore


Sr. No	Item Name	Model	Make	Status	Observations raised by Technical Committee	Reply by the firm	Review by the Technical Member	Decision of the Committee	Current Status
1	Electrohydraulic OT Table	Kepler T3000	Medicinos	Non Responsive	Electric lateral tilt 20° and -20° instead of 30° degree and -30° degree or	Reply submitted by firm	Objection Removed	Grievance Accepted, Objection removed	Responsive



Muhammad Salman Khan
Procurement Officer
SZMC/SZH, R. Y. Khan
(Member)



Engr. Awais Bin Tariq
Bio Medical Engineer)
SZMC/SZH, R. Y. Khan
(Technical (Member)



Dr. Saira Sadaf
HOD of Anesthesia Department
SZMC/SZH, R. Y. Khan
(Member)



Dr. Mushtaq Ahmed Azmi
Director Emergency)
SZMC/SZH, R. Y. Khan
(Member)


Prof. Dr. Shazia Majid Khan
HOD Gynae/Obst Department
SZMC/SZH, R. Y. Khan
(Member)


Dr. Yaseen Malik
AMS
SZH, R. Y. Khan (Member)


Dr. Moeen Akhter Malik
Assistant Prof. Medicine Department
SZMC/SZH, R. Y. Khan (Member)


Dr. Ali Burhan Mustafa
Associate Prof./HOD Psychiatry Department
SZMC/SZH, R. Y. Khan (Member)


Dr. Muhammad Tariq Ghafoor
Associate Professor /HOD Surgery Department
(Chairman) SZMC/SZH, R. Y. Khan